


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 4372 Accredited to ISO/IEC 17025:2005	Intech Calibration Ltd	
	Issue No: 004 Issue date: 06 October 2010	
	Unit 8, The Azure Suite, Churchill Court, 112 The Street, Rustington West Sussex BN16 3DA	Contact: Mr J du Plessis Tel: +44 (0)1903 772 859 Fax: +44 (0)1903 754 437 E-Mail: info@intechcalibration.co.uk Website: www.intechcalibration.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories and Manufacturing & Processing Plants	Temperature	S



4372
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Intech Calibration Limited
Issue No: 004 Issue date: 06 October 2010

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
TEMPERATURE Temperature indicators with sensors	- 90 °C to 125 °C	0.15 °C	Sensors with suitable dimensions	S
END				